
Midmark M11 - v1.0.5
Total Cycles: 810

Name: RITTER2
Sterilizer ID: V2313228
02/16/2022 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.6 F 0.0
2:00 155.5 F 2.9
4:00 172.9 F 6.8
6:00 197.1 F 11.9
8:00 216.3 F 6.0
10:00 229.9 F 11.5
12:00 233.8 F 9.0
14:00 238.6 F 9.9
16:00 248.0 F 14.7
18:00 260.3 F 21.7
20:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.7 F 29.8
2:00 271.5 F 29.0
3:00 272.8 F 30.0
4:00 272.3 F 29.4

Min 271.4 F 29.0
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 210.7 F 0.0
10:00 219.4 F 0.0
15:00 217.9 F 0.0
20:00 223.6 F 0.0
25:00 221.8 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:36
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

02/16/2022 03:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____