
Midmark M11 - v1.0.5
Total Cycles: 3387

Name: RITTER2
Sterilizer ID: V2313228
02/15/2024 11:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.5 F 0.0
2:00 130.7 F 0.2
4:00 154.4 F 2.2
6:00 170.1 F 5.5
8:00 191.4 F 9.9
10:00 206.5 F 6.2
12:00 217.1 F 7.5
14:00 232.7 F 12.4
16:00 234.2 F 9.5
18:00 237.5 F 9.5
20:00 247.2 F 14.3
22:00 259.1 F 20.8
24:00 269.5 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 30.0
2:00 272.4 F 29.5
3:00 271.4 F 28.9
4:00 271.7 F 29.5

Min 271.2 F 28.9
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 191.9 F 0.0
10:00 190.4 F 0.0
15:00 190.0 F 0.0
20:00 187.9 F 0.0
25:00 186.4 F 0.0
30:00 186.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 25:05
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:03:19
CYCLE COMPLETE

02/15/2024 12:38 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____