
Midmark M11 - v1.0.5
Total Cycles: 2088

Name: RITTER2
Sterilizer ID: V2313228
02/23/2023 04:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.0 F 0.0
2:00 179.3 F 2.7
4:00 199.4 F 7.4
6:00 218.5 F 13.2
8:00 228.1 F 9.2
10:00 235.9 F 9.4
12:00 243.7 F 12.8
14:00 245.2 F 13.4
16:00 258.7 F 20.7
18:00 270.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.5
2:00 271.3 F 29.1
3:00 272.5 F 29.6
4:00 271.4 F 28.9

Min 271.2 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 211.8 F 0.0
10:00 227.3 F 0.0
15:00 230.0 F 0.0
20:00 227.8 F 0.0
25:00 227.6 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 18:45
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:35
CYCLE COMPLETE

02/23/2023 05:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____