
Midmark M11 - v1.0.5
Total Cycles: 2847

Name: RITTER2
Sterilizer ID: V2313228
09/07/2023 04:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.1 F 0.0
2:00 168.4 F 3.0
4:00 185.3 F 7.2
6:00 208.1 F 12.0
8:00 219.0 F 6.4
10:00 230.7 F 11.2
12:00 231.7 F 8.3
14:00 240.6 F 11.5
16:00 245.0 F 13.3
18:00 257.1 F 19.8
20:00 267.9 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.7 F 29.8
2:00 271.5 F 29.0
3:00 272.6 F 29.9
4:00 272.2 F 29.4

Min 271.3 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 203.0 F 0.0
10:00 210.3 F 0.0
15:00 214.6 F 0.0
20:00 213.4 F 0.0
25:00 210.2 F 0.0
30:00 208.7 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

09/07/2023 05:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____