

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4829

Name: RITTER 1  
Sterilizer ID: V2178198  
08/19/2024 08:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.1 F 0.0  
2:00 68.7 F 0.0  
4:00 82.7 F 0.1  
6:00 116.9 F 0.3  
8:00 146.5 F 0.5  
10:00 170.6 F 0.7  
12:00 196.6 F 1.2  
14:00 210.0 F 3.6  
16:00 219.1 F 6.7  
18:00 228.4 F 10.1  
20:00 235.3 F 9.7  
22:00 238.0 F 10.3  
24:00 236.3 F 8.5  
26:00 245.9 F 12.6  
28:00 245.0 F 12.6  
30:00 257.4 F 19.1  
32:00 267.9 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 272.5 F 28.8  
2:00 272.6 F 28.6  
3:00 271.8 F 28.1  
4:00 271.9 F 28.4  
  
Min 271.6 F 28.0  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 198.1 F 0.0  
10:00 206.0 F 0.0  
15:00 212.5 F 0.0  
20:00 213.8 F 0.0  
25:00 217.9 F 0.0  
30:00 227.9 F 0.0  
  
Drying Complete

FILLING: 0:40  
HEATING: 33:28  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:11:34  
CYCLE COMPLETE

08/19/2024 09:46 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_