

-----  
Midmark M11 - v1.0.0  
Total Cycles: 620

Name: MIDMARK  
Sterilizer ID: V2178198  
01/21/2021 03:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
2:00 198.5 F 2.9  
4:00 193.8 F 7.3  
6:00 215.6 F 12.9  
8:00 222.9 F 8.6  
10:00 229.1 F 10.4  
12:00 241.1 F 11.9  
14:00 239.5 F 10.4  
16:00 253.4 F 17.2  
18:00 265.1 F 24.1  
20:00 271.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.1  
1:00 272.4 F 29.4  
2:00 272.1 F 29.1  
3:00 271.7 F 29.2  
4:00 272.2 F 29.5

Min 271.5 F 29.0  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
5:00 230.9 F 0.0  
10:00 226.2 F 0.0  
15:00 227.7 F 0.0  
20:00 226.1 F 0.0  
25:00 220.6 F 0.0  
Dry Cycle Aborted

FILLING: 1:37  
HEATING: 20:23  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 27:04  
TOTAL CYCLE: 00:56:19

01/21/2021 04:49 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_