
Midmark M11 - v1.0.0
Total Cycles: 2976

Name: RITTER 1
Sterilizer ID: V2178198
03/08/2023 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.5 F 0.0
2:00 77.7 F 0.0
4:00 98.9 F 0.0
6:00 119.4 F 0.0
8:00 158.9 F 0.1
10:00 182.4 F 2.6
12:00 197.0 F 6.0
14:00 214.3 F 9.2
16:00 225.4 F 12.8
18:00 228.2 F 6.7
20:00 237.2 F 10.9
22:00 233.5 F 7.5
24:00 245.1 F 12.8
26:00 242.3 F 11.1
28:00 254.7 F 17.4
30:00 265.4 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 272.7 F 28.6
2:00 271.3 F 27.9
3:00 272.8 F 28.7
4:00 271.7 F 27.9

Min 271.2 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 238.4 F 0.0
10:00 209.2 F 0.0
15:00 224.3 F 0.0
20:00 226.7 F 0.0
25:00 232.6 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:54
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:50
CYCLE COMPLETE

03/08/2023 09:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____