
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000483
12:15 19/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:19
STERILIZING 4:28
136.0 C 228kPa 4:28
136.4 C 230kPa 4:58
136.4 C 233kPa 5:28
136.3 C 232kPa 5:58
136.4 C 233kPa 6:28
136.4 C 233kPa 6:58
136.3 C 230kPa 7:28
136.3 C 232kPa 7:58
136.3 C 232kPa 7:58
VENTING 7:58
AIR DRYING 8:53
CYCLE COMPLETE 23:53

Digital Signature #
6EE2C8615E727EB4
