

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4460

Name: RITTER 1  
Sterilizer ID: V2178198  
04/15/2024 04:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 109.9 F 0.0  
2:00 111.6 F 1.1  
4:00 123.0 F 3.4  
6:00 179.1 F 6.2  
8:00 190.9 F 10.1  
10:00 211.2 F 8.2  
12:00 214.0 F 5.8  
14:00 224.3 F 9.7  
16:00 233.8 F 10.5  
18:00 236.5 F 10.4  
20:00 238.7 F 8.4  
22:00 246.8 F 13.9  
24:00 257.8 F 19.7  
26:00 267.9 F 26.0  
28:00 272.2 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.4 F 29.2  
2:00 272.5 F 29.6  
3:00 271.7 F 28.9  
4:00 272.4 F 29.4

Min 271.5 F 28.9  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 218.4 F 0.0  
10:00 216.5 F 0.0  
15:00 212.4 F 0.0  
20:00 216.9 F 0.0  
25:00 224.7 F 0.0  
30:00 229.3 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 28:08  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:07:03  
CYCLE COMPLETE

04/15/2024 05:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_