
Midmark M11 - v1.0.5
Total Cycles: 946

Name: RITTER2
Sterilizer ID: V2313228
04/01/2022 01:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.9 F 0.0
2:00 196.1 F 2.1
4:00 206.4 F 7.6
6:00 224.7 F 13.3
8:00 231.6 F 10.1
10:00 233.8 F 8.5
12:00 239.1 F 10.4
14:00 248.5 F 15.3
16:00 261.6 F 22.7
18:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.4 F 29.6
2:00 272.5 F 30.1
3:00 271.8 F 29.3
4:00 272.7 F 30.1

Min 271.6 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 213.5 F 0.0
10:00 237.7 F 0.0
15:00 240.5 F 0.0
20:00 227.8 F 0.0
25:00 216.0 F 0.0
30:00 213.8 F 0.0
Drying Complete

FILLING: 2:47
HEATING: 18:14
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

04/01/2022 02:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____