
Midmark M11 - v1.0.5
Total Cycles: 2047

Name: RITTER2
Sterilizer ID: V2313228
02/14/2023 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.2 F 0.0
2:00 179.7 F 2.7
4:00 187.1 F 6.7
6:00 208.4 F 11.9
8:00 221.3 F 7.1
10:00 234.5 F 12.6
12:00 237.5 F 10.4
14:00 236.1 F 9.1
16:00 250.7 F 16.1
18:00 262.5 F 22.9
20:00 272.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.2 F 29.1
2:00 272.8 F 29.6
3:00 271.7 F 29.3
4:00 272.1 F 29.2

Min 271.5 F 28.9
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 221.5 F 0.0
10:00 214.2 F 0.0
15:00 216.5 F 0.0
20:00 217.9 F 0.0
25:00 218.5 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:47
CYCLE COMPLETE

02/14/2023 01:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____