
Midmark M11 - v1.0.5
Total Cycles: 28

Name: RITTER2
Sterilizer ID: V2313228
05/19/2021 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 71.5 F 0.3
4:00 76.3 F 1.1
6:00 101.6 F 3.1
8:00 142.2 F 5.9
10:00 164.7 F 9.1
12:00 185.7 F 12.7
14:00 209.1 F 4.7
16:00 221.4 F 8.8
18:00 232.8 F 13.2
20:00 234.7 F 9.3
22:00 240.1 F 10.6
24:00 244.9 F 12.9
26:00 256.5 F 19.2
28:00 267.2 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.7 F 29.3
2:00 272.0 F 29.2
3:00 272.3 F 29.0
4:00 272.1 F 29.5

Min 271.5 F 28.6
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.3 F 0.0
5:00 213.2 F 0.0
10:00 207.3 F 0.0
15:00 203.9 F 0.0
20:00 210.4 F 0.0
25:00 207.8 F 0.0
30:00 208.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 29:28
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:07:29
CYCLE COMPLETE

05/19/2021 09:09 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____