
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002212
11:30 10/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:33
STERILIZING 4:34
136.4 C 220kPa 4:34
136.1 C 222kPa 5:05
136.2 C 224kPa 5:35
136.4 C 224kPa 6:05
136.2 C 221kPa 6:35
136.4 C 223kPa 7:05
136.4 C 224kPa 7:35
136.4 C 224kPa 8:05
136.4 C 224kPa 8:05
VENTING 8:05
AIR DRYING 8:59
CYCLE COMPLETE 9:04

Digital Signature #
FC65A30CC1F51908
