
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000871
16:22 14/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:18
STERILIZING 3:25
136.0 C 220kPa 3:25
136.1 C 229kPa 3:55
136.4 C 232kPa 4:25
136.2 C 231kPa 4:55
136.2 C 231kPa 5:25
136.1 C 232kPa 5:55
136.4 C 233kPa 6:25
136.4 C 232kPa 6:55
136.4 C 232kPa 6:55
VENTING 6:55
AIR DRYING 7:47
CYCLE COMPLETE 22:47

Digital Signature #
4B9C6AE3264F08E9
