
Midmark M11 - v1.0.0
Total Cycles: 1016

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2021 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.1 F 0.0
2:00 185.2 F 2.7
4:00 193.1 F 6.5
6:00 210.5 F 11.5
8:00 228.7 F 6.9
10:00 239.2 F 12.4
12:00 238.6 F 9.6
14:00 237.8 F 8.9
16:00 248.6 F 14.2
18:00 260.3 F 20.7
20:00 270.9 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.0 F 28.4
2:00 271.9 F 28.7
3:00 272.3 F 29.0
4:00 271.5 F 28.5

Min 271.5 F 28.3
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 236.1 F 0.0
10:00 222.8 F 0.0
15:00 218.6 F 0.0
20:00 211.5 F 0.0
25:00 208.8 F 0.0
30:00 206.4 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

07/15/2021 10:00 AM
ACCEPTED: X
REJECTED:
Operator: U02