
Midmark M11 - v1.0.5
Total Cycles: 3161

Name: RITTER2
Sterilizer ID: V2313228
11/27/2023 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.0 F 0.0
2:00 118.1 F 0.2
4:00 132.7 F 1.7
6:00 163.2 F 5.1
8:00 184.5 F 9.4
10:00 197.9 F 9.4
12:00 216.7 F 6.5
14:00 228.4 F 11.0
16:00 229.5 F 7.6
18:00 242.9 F 13.4
20:00 242.4 F 12.2
22:00 254.8 F 18.6
24:00 266.0 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.8 F 29.8
2:00 271.5 F 29.1
3:00 272.9 F 29.9
4:00 271.8 F 29.1

Min 271.5 F 29.0
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.0 F 0.0
5:00 199.3 F 0.0
10:00 194.7 F 0.0
15:00 194.9 F 0.0
20:00 192.7 F 0.0
25:00 190.4 F 0.0
30:00 189.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:40
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:03:53
CYCLE COMPLETE

11/27/2023 12:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____