
Midmark M11 - v1.0.5
Total Cycles: 659

Name: RITTER2
Sterilizer ID: V2313228
12/21/2021 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.6 F 0.0
2:00 197.8 F 3.8
4:00 201.6 F 9.3
6:00 220.7 F 7.3
8:00 232.6 F 10.9
10:00 233.6 F 8.8
12:00 239.2 F 10.1
14:00 248.3 F 15.0
16:00 261.0 F 22.1
18:00 272.2 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.9
2:00 272.0 F 29.3
3:00 271.4 F 29.1
4:00 272.6 F 30.0

Min 271.4 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 217.1 F 0.0
10:00 234.2 F 0.0
15:00 233.7 F 0.0
20:00 226.4 F 0.0
25:00 221.7 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 2:39
HEATING: 18:32
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

12/21/2021 01:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____