

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4065

Name: RITTER2  
Sterilizer ID: V2313228  
10/29/2024 04:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.5 F 0.0  
2:00 163.6 F 3.0  
4:00 190.6 F 8.1  
6:00 209.7 F 10.1  
8:00 227.4 F 9.1  
10:00 233.2 F 8.8  
12:00 243.2 F 13.4  
14:00 243.8 F 12.6  
16:00 257.0 F 19.7  
18:00 268.9 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.2 F 29.3  
2:00 272.4 F 29.8  
3:00 271.3 F 29.2  
4:00 272.6 F 29.5  
  
Min 271.3 F 29.0  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.6 F 0.0  
5:00 211.6 F 0.0  
10:00 203.9 F 0.0  
15:00 200.5 F 0.0  
20:00 199.4 F 0.0  
25:00 202.1 F 0.0  
30:00 203.1 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 19:02  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 00:58:04  
CYCLE COMPLETE

10/29/2024 05:35 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_