
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002394
8:31 08/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:16
STERILIZING 4:22
136.5 C 223kPa 4:22
136.2 C 223kPa 4:52
136.2 C 219kPa 5:22
136.2 C 221kPa 5:52
136.4 C 223kPa 6:22
136.4 C 224kPa 6:52
136.4 C 223kPa 7:22
136.3 C 221kPa 7:52
136.3 C 221kPa 7:52
VENTING 7:52
AIR DRYING 8:42
CYCLE COMPLETE 8:47

Digital Signature #
3A45C023792811B9
