

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3611

Name: RITTER 1  
Sterilizer ID: V2178198  
08/03/2023 11:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.8 F 0.0  
2:00 127.4 F 0.0  
4:00 151.9 F 0.0  
6:00 190.9 F 3.6  
8:00 199.2 F 8.4  
10:00 215.5 F 13.1  
12:00 225.1 F 7.4  
14:00 237.2 F 12.2  
16:00 235.6 F 9.2  
18:00 238.7 F 9.6  
20:00 247.1 F 13.9  
22:00 259.3 F 20.5  
24:00 270.3 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.5 F 28.9  
2:00 271.9 F 28.7  
3:00 272.4 F 28.9  
4:00 271.7 F 28.6

Min 271.6 F 28.3  
Max 273.0 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.5 F 0.0  
5:00 212.5 F 0.0  
10:00 229.8 F 0.0  
15:00 226.6 F 0.0  
20:00 227.3 F 0.0  
25:00 228.7 F 0.0  
30:00 227.4 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 24:50  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:02:55  
CYCLE COMPLETE

08/03/2023 12:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_