
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000217
10:56 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:09
PRESSURIZING 4:10
STERILIZING 5:40
136.4 C 233kPa 5:40
136.4 C 233kPa 6:10
136.3 C 232kPa 6:40
136.4 C 233kPa 7:10
136.4 C 233kPa 7:40
136.4 C 233kPa 8:10
136.4 C 233kPa 8:40
136.4 C 233kPa 9:10
136.4 C 233kPa 9:10
VENTING 9:10
AIR DRYING 10:15
CYCLE COMPLETE 25:15

Digital Signature #
589A527D91E45D3D
