
Midmark M11 - v1.0.0
Total Cycles: 3433

Name: RITTER 1
Sterilizer ID: V2178198
06/23/2023 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.7 F 0.0
2:00 181.3 F 1.0
4:00 192.7 F 4.3
6:00 213.6 F 9.6
8:00 223.0 F 7.3
10:00 232.6 F 9.9
12:00 235.0 F 8.8
14:00 244.1 F 12.9
16:00 242.3 F 11.4
18:00 254.9 F 17.8
20:00 265.8 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.3 F 29.0
2:00 271.9 F 28.4
3:00 272.0 F 28.8
4:00 272.2 F 28.7

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 211.8 F 0.0
10:00 228.8 F 0.0
15:00 233.0 F 0.0
20:00 223.6 F 0.0
25:00 220.4 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 21:37
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:30
CYCLE COMPLETE

06/23/2023 09:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____