
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002589
14:14 21/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:07
STERILIZING 3:05
136.4 C 223kPa 3:05
136.4 C 226kPa 3:35
136.4 C 226kPa 4:05
136.3 C 223kPa 4:35
136.3 C 224kPa 5:05
136.4 C 226kPa 5:35
136.4 C 225kPa 6:05
136.4 C 224kPa 6:35
136.4 C 224kPa 6:35
VENTING 6:35
AIR DRYING 7:27
CYCLE COMPLETE 14:25

Digital Signature #
63D8DA6F7CBF6DA8
