
Midmark M11 - v1.0.0
Total Cycles: 5033

Name: RITTER 1
Sterilizer ID: V2178198
10/29/2024 11:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.4 F 0.0
2:00 156.0 F 0.3
4:00 177.9 F 1.8
6:00 199.6 F 6.2
8:00 215.2 F 10.7
10:00 225.0 F 5.5
12:00 231.5 F 10.6
14:00 233.4 F 7.3
16:00 244.8 F 12.7
18:00 243.8 F 12.4
20:00 256.9 F 19.3
22:00 268.5 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.6 F 29.0
2:00 271.8 F 28.8
3:00 272.6 F 29.0
4:00 271.5 F 28.5

Min 271.5 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.8 F 0.0
5:00 226.0 F 0.0
10:00 235.7 F 0.0
15:00 227.9 F 0.0
20:00 223.1 F 0.0
25:00 221.0 F 0.0
30:00 211.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 23:09
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:36
CYCLE COMPLETE

10/29/2024 12:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____