
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000086
15:38 21/08/2020
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 4:38
STERILIZING 5:36
136.4 C 226kPa 5:36
136.2 C 223kPa 6:06
136.4 C 224kPa 6:36
136.1 C 222kPa 7:06
136.0 C 221kPa 7:36
136.4 C 226kPa 8:06
136.1 C 223kPa 8:36
136.4 C 225kPa 9:06
136.3 C 225kPa 9:36
136.4 C 225kPa 10:06
136.3 C 224kPa 10:36
136.2 C 224kPa 11:06
136.4 C 224kPa 11:36
136.4 C 225kPa 12:06
136.4 C 225kPa 12:36
136.2 C 223kPa 13:06
136.4 C 225kPa 13:36
136.4 C 225kPa 14:06
136.1 C 223kPa 14:36
136.4 C 226kPa 15:06
136.4 C 226kPa 15:36
136.4 C 226kPa 15:36
VENTING 15:37
AIR DRYING 16:22
CYCLE COMPLETE 25:54

Digital Signature #
B14ED48A714F67CA
