
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000617
8:38 14/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:18
STERILIZING 4:23
136.0 C 220kPa 4:23
135.9 C 217kPa 4:53
135.9 C 220kPa 5:23
136.2 C 224kPa 5:53
136.3 C 222kPa 6:23
135.9 C 220kPa 6:53
136.4 C 224kPa 7:23
136.3 C 223kPa 7:53
136.3 C 223kPa 7:53
VENTING 7:53
AIR DRYING 8:44
CYCLE COMPLETE 8:54

Digital Signature #
5F7144F122A3C9EB
