
Midmark M11 - v1.0.5
Total Cycles: 766

Name: RITTER2
Sterilizer ID: V2313228
02/07/2022 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.0 F 0.0
2:00 153.6 F 2.5
4:00 176.4 F 6.6
6:00 200.7 F 11.5
8:00 217.0 F 5.4
10:00 230.5 F 10.5
12:00 232.6 F 8.0
14:00 241.6 F 11.5
16:00 246.1 F 13.7
18:00 258.6 F 20.5
20:00 269.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.7 F 29.8
2:00 272.2 F 29.1
3:00 271.4 F 29.0
4:00 272.8 F 29.8

Min 271.4 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.7 F 0.0
5:00 222.0 F 0.0
10:00 205.8 F 0.0
15:00 206.3 F 0.0
20:00 203.3 F 0.0
25:00 204.0 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:60
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:30
CYCLE COMPLETE

02/07/2022 04:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____