
Midmark M11 - v1.0.0
Total Cycles: 2367

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2022 09:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.8 F 0.0
2:00 127.8 F 1.1
4:00 148.3 F 4.0
6:00 167.7 F 7.8
8:00 190.6 F 12.2
10:00 208.3 F 4.2
12:00 220.3 F 8.8
14:00 231.9 F 13.5
16:00 235.4 F 9.7
18:00 236.9 F 9.1
20:00 246.1 F 13.6
22:00 257.7 F 19.9
24:00 268.7 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.5 F 29.4
2:00 271.5 F 28.7
3:00 272.5 F 29.5
4:00 272.0 F 28.9

Min 271.5 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 219.0 F 0.0
10:00 229.6 F 0.0
15:00 229.5 F 0.0
20:00 223.8 F 0.0
25:00 225.0 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:12
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:03:18
CYCLE COMPLETE

09/26/2022 10:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____