
Midmark M11 - v1.0.0
Total Cycles: 5404

Name: RITTER 1
Sterilizer ID: V2178198
03/10/2025 02:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.2 F 0.0
2:00 94.9 F 0.7
4:00 101.4 F 2.7
6:00 143.1 F 5.5
8:00 170.0 F 8.4
10:00 196.9 F 12.1
12:00 208.8 F 2.7
14:00 216.3 F 5.9
16:00 226.8 F 9.3
18:00 235.9 F 13.3
20:00 236.6 F 9.2
22:00 241.7 F 11.0
24:00 244.4 F 12.3
26:00 255.3 F 18.0
28:00 265.4 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 271.8 F 28.9
2:00 272.4 F 29.2
3:00 271.6 F 28.7
4:00 272.5 F 29.4

Min 270.9 F 28.2
Max 273.2 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 204.2 F 0.0
10:00 220.4 F 0.0
15:00 248.0 F 0.0
20:00 229.6 F 0.0
25:00 226.7 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:49
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:08:35
CYCLE COMPLETE

03/10/2025 03:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____