
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000816
12:09 25/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:15
PRESSURIZING 5:56
STERILIZING 7:22
136.0 C 220kPa 7:22
136.4 C 224kPa 7:52
136.0 C 220kPa 8:22
136.3 C 223kPa 8:52
136.3 C 222kPa 9:22
136.2 C 223kPa 9:52
136.2 C 220kPa 10:22
136.4 C 223kPa 10:52
136.4 C 223kPa 10:52
VENTING 10:53
AIR DRYING 11:51
CYCLE COMPLETE 41:51

Digital Signature #
41A99A7FAD4363DC
