
Midmark M11 - v1.0.5
Total Cycles: 4329

Name: RITTER2
Sterilizer ID: V2313228
02/03/2025 03:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.6 F 0.0
2:00 161.6 F 3.0
4:00 172.3 F 6.8
6:00 199.4 F 11.9
8:00 214.9 F 5.4
10:00 228.8 F 10.2
12:00 231.9 F 8.4
14:00 242.7 F 13.1
16:00 242.3 F 12.1
18:00 255.0 F 18.6
20:00 266.2 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.9 F 29.7
2:00 271.8 F 29.2
3:00 272.9 F 29.8
4:00 271.8 F 29.0

Min 271.7 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 178.8 F 0.0
10:00 203.9 F 0.0
15:00 220.8 F 0.0
Dry Cycle Aborted

FILLING: 1:04
HEATING: 21:37
STERILIZING: 4:00
VENTING: 3:27
DRYING: 19:07
TOTAL CYCLE: 00:49:15

02/03/2025 04:05 PM
ACCEPTED:
REJECTED:
Reason: -----
Operator: -----