

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001039  
15:26 13/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:44  
PRESSURIZING 2:55  
STERILIZING 4:18  
136.0 C 220kPa 4:18  
136.4 C 225kPa 4:48  
136.2 C 223kPa 5:18  
136.4 C 225kPa 5:48  
136.3 C 225kPa 6:18  
136.3 C 223kPa 6:48  
136.4 C 225kPa 7:18  
136.4 C 226kPa 7:48  
136.4 C 226kPa 7:48  
VENTING 7:49  
AIR DRYING 8:47  
CYCLE COMPLETE 12:31  
-----

Digital Signature #  
2464E545A84CC5EF  
-----