

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 003379  
12:26 04/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 0:41  
PRESSURIZING 0:50  
STERILIZING 1:55  
136.0 C 220kPa 1:55  
136.0 C 222kPa 2:25  
136.4 C 224kPa 2:55  
136.1 C 224kPa 3:25  
136.4 C 223kPa 3:55  
136.1 C 219kPa 4:25  
136.4 C 224kPa 4:55  
136.2 C 221kPa 5:25  
136.2 C 221kPa 5:25  
VENTING 5:25  
AIR DRYING 6:18  
CYCLE COMPLETE 7:58  
-----

Digital Signature #  
B9E7FBDA9D89FFA1  
-----