
Midmark M11 - v1.0.0
Total Cycles: 1592

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2022 04:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.0 F 0.0
2:00 194.1 F 4.4
4:00 215.0 F 10.0
6:00 222.3 F 5.7
8:00 237.5 F 11.4
10:00 235.6 F 9.0
12:00 239.7 F 9.3
14:00 249.5 F 14.9
16:00 262.3 F 22.1
18:00 272.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.7
2:00 271.6 F 28.4
3:00 272.5 F 28.7
4:00 271.5 F 28.1

Min 271.4 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 226.7 F 0.0
10:00 221.4 F 0.0
15:00 221.8 F 0.0
20:00 223.0 F 0.0
25:00 229.8 F 0.0
30:00 232.3 F 0.0
Drying Complete

FILLING: 2:29
HEATING: 18:12
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

02/02/2022 05:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____