

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2113

Name: RITTER 1  
Sterilizer ID: V2178198  
07/11/2022 12:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.3 F 0.0  
2:00 131.6 F 0.2  
4:00 172.3 F 1.9  
6:00 176.9 F 5.0  
8:00 196.6 F 8.8  
10:00 214.9 F 13.3  
12:00 225.4 F 7.0  
14:00 237.4 F 11.4  
16:00 232.8 F 7.9  
18:00 245.1 F 13.5  
20:00 242.8 F 11.7  
22:00 254.8 F 17.9  
24:00 265.5 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 29.0  
2:00 271.5 F 28.2  
3:00 272.1 F 29.0  
4:00 272.3 F 28.8

Min 271.3 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.6 F 0.0  
5:00 224.2 F 0.0  
10:00 216.8 F 0.0  
15:00 219.8 F 0.0  
20:00 234.8 F 0.0  
25:00 233.9 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 25:53  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:03:54  
CYCLE COMPLETE

07/11/2022 01:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_