
Midmark M11 - v1.0.5
Total Cycles: 1037

Name: RITTER2
Sterilizer ID: V2313228
04/26/2022 02:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.0 F 0.0
2:00 160.0 F 2.9
4:00 179.9 F 6.9
6:00 202.5 F 11.9
8:00 220.7 F 6.2
10:00 228.8 F 11.3
12:00 233.0 F 8.4
14:00 240.5 F 11.2
16:00 245.3 F 13.4
18:00 257.6 F 20.0
20:00 268.6 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 271.5 F 29.2
2:00 272.8 F 30.0
3:00 273.0 F 29.6
4:00 272.3 F 29.3

Min 271.1 F 28.8
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 240.5 F 0.0
10:00 226.2 F 0.0
15:00 229.4 F 0.0
20:00 206.4 F 0.0
25:00 187.8 F 0.0
30:00 184.2 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:06
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

04/26/2022 03:49 PM
ACCEPTED: X
REJECTED:
Operator: U02