
Midmark M11 - v1.0.5
Total Cycles: 797

Name: RITTER2
Sterilizer ID: V2313228
02/14/2022 03:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.2 F 0.0
2:00 165.1 F 3.6
4:00 185.2 F 8.3
6:00 211.1 F 10.3
8:00 225.9 F 9.1
10:00 234.8 F 9.5
12:00 242.9 F 13.2
14:00 243.7 F 12.6
16:00 257.1 F 19.7
18:00 269.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.2 F 29.2
2:00 272.6 F 29.7
3:00 271.7 F 28.9
4:00 272.4 F 29.6

Min 271.5 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.1 F 0.0
5:00 217.6 F 0.0
10:00 228.2 F 0.0
15:00 225.7 F 0.0
20:00 230.1 F 0.0
25:00 224.8 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:01
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

02/14/2022 04:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____