
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001600
10:57 16/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:03
STERILIZING 3:04
136.4 C 224kPa 3:04
136.4 C 226kPa 3:34
136.4 C 225kPa 4:04
136.4 C 225kPa 4:34
136.4 C 225kPa 5:04
136.4 C 225kPa 5:34
136.4 C 226kPa 6:04
136.4 C 225kPa 6:34
136.4 C 225kPa 6:34
VENTING 6:34
AIR DRYING 7:27
CYCLE COMPLETE 7:34

Digital Signature #
CB35EDA80783625C
