
Midmark M11 - v1.0.5
Total Cycles: 3663

Name: RITTER2
Sterilizer ID: V2313228
06/03/2024 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.8 F 0.0
2:00 101.8 F 0.4
4:00 122.6 F 2.2
6:00 152.6 F 4.7
8:00 174.7 F 7.8
10:00 191.4 F 11.1
12:00 209.7 F 4.5
14:00 215.9 F 7.0
16:00 226.2 F 10.9
18:00 230.2 F 7.9
20:00 238.1 F 11.6
22:00 236.1 F 9.3
24:00 248.1 F 14.9
26:00 258.1 F 20.5
28:00 267.6 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 29.9
2:00 271.5 F 29.1
3:00 272.2 F 30.0
4:00 272.1 F 29.5

Min 271.3 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 198.5 F 0.0
10:00 203.9 F 0.0
15:00 197.5 F 0.0
20:00 196.1 F 0.0
25:00 198.0 F 0.0
30:00 198.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:32
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:07:59
CYCLE COMPLETE

06/03/2024 02:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____