

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4392

Name: RITTER2  
Sterilizer ID: V2313228  
02/28/2025 03:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.0 F 0.0  
2:00 126.0 F 1.4  
4:00 131.1 F 3.8  
6:00 158.8 F 7.5  
8:00 189.5 F 11.7  
10:00 209.2 F 5.6  
12:00 221.4 F 10.5  
14:00 228.2 F 7.5  
16:00 238.9 F 12.6  
18:00 239.6 F 11.0  
20:00 251.6 F 17.3  
22:00 262.2 F 23.5  
24:00 272.4 F 30.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 30.0  
1:00 271.5 F 29.9  
2:00 272.0 F 30.0  
3:00 272.4 F 30.3  
4:00 272.0 F 30.5

Min 271.2 F 29.9  
Max 272.8 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.6 F 0.0  
5:00 198.1 F 0.0  
10:00 203.4 F 0.0  
15:00 203.8 F 0.0  
20:00 199.5 F 0.0  
25:00 198.2 F 0.0  
30:00 197.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 24:23  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:03:03  
CYCLE COMPLETE

02/28/2025 04:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_