
Midmark M11 - v1.0.0
Total Cycles: 904

Name: RITTER 1
Sterilizer ID: V2178198
06/16/2021 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.8 F 0.0
2:00 163.2 F 0.0
4:00 147.1 F 1.5
6:00 164.3 F 4.0
8:00 197.6 F 7.5
10:00 217.1 F 11.7
12:00 224.2 F 5.8
14:00 229.9 F 10.3
16:00 235.7 F 8.4
18:00 242.1 F 11.9
20:00 238.4 F 9.7
22:00 250.9 F 15.7
24:00 261.9 F 22.0
26:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.8 F 29.1
2:00 271.7 F 28.9
3:00 272.7 F 29.0
4:00 271.5 F 28.8

Min 271.4 F 28.6
Max 273.2 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 237.2 F 0.0
10:00 225.4 F 0.0
15:00 231.6 F 0.0
20:00 229.3 F 0.0
25:00 197.0 F 0.0
30:00 195.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 27:17
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:05:35
CYCLE COMPLETE

06/16/2021 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____