
Midmark M11 - v1.0.5
Total Cycles: 3516

Name: RITTER2
Sterilizer ID: V2313228
04/11/2024 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.9 F 0.0
2:00 129.5 F 0.2
4:00 147.7 F 2.4
6:00 172.4 F 5.6
8:00 194.9 F 9.7
10:00 211.3 F 9.6
12:00 218.8 F 7.6
14:00 232.1 F 12.0
16:00 233.8 F 8.8
18:00 240.9 F 11.3
20:00 244.5 F 12.8
22:00 255.6 F 18.9
24:00 265.9 F 25.2
26:00 272.1 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.6 F 29.7
2:00 271.7 F 29.7
3:00 272.3 F 29.6
4:00 271.6 F 29.7

Min 271.3 F 29.2
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.5 F 0.0
5:00 232.0 F 0.0
10:00 217.3 F 0.0
15:00 188.6 F 0.0
20:00 184.3 F 0.0
25:00 182.9 F 0.0
30:00 183.9 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:32
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:05:08
CYCLE COMPLETE

04/11/2024 11:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____