
Midmark M11 - v1.0.5
Total Cycles: 1380

Name: RITTER2
Sterilizer ID: V2313228
08/09/2022 07:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 78.4 F 0.0
4:00 84.8 F 0.0
6:00 103.7 F 0.1
8:00 136.6 F 2.7
10:00 167.2 F 6.0
12:00 206.5 F 9.6
14:00 220.0 F 10.5
16:00 222.1 F 7.2
18:00 232.9 F 11.3
20:00 232.8 F 7.8
22:00 242.7 F 13.3
24:00 242.3 F 11.6
26:00 253.8 F 17.6
28:00 264.4 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 271.9 F 29.8
2:00 271.8 F 30.0
3:00 271.9 F 30.0
4:00 271.9 F 29.9

Min 270.0 F 28.4
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 206.1 F 0.0
10:00 216.5 F 0.0
15:00 223.6 F 0.0
20:00 226.6 F 0.0
25:00 228.8 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:57
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:07:55
CYCLE COMPLETE

08/09/2022 08:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____