
Midmark M11 - v1.0.0
Total Cycles: 4201

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2024 05:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.7 F 0.0
2:00 119.6 F 1.3
4:00 132.3 F 4.0
6:00 162.0 F 7.4
8:00 182.5 F 11.5
10:00 209.4 F 3.9
12:00 218.2 F 8.0
14:00 228.9 F 12.3
16:00 231.7 F 8.2
18:00 243.3 F 13.4
20:00 241.7 F 11.4
22:00 253.4 F 17.3
24:00 264.2 F 23.6
26:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.2
2:00 271.6 F 28.5
3:00 272.2 F 29.3
4:00 272.3 F 29.0

Min 271.3 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 204.1 F 0.0
10:00 208.1 F 0.0
15:00 210.3 F 0.0
20:00 215.3 F 0.0
25:00 223.3 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 26:07
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:04:44
CYCLE COMPLETE

01/12/2024 06:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____