

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2378

Name: RITTER2  
Sterilizer ID: V2313228  
05/16/2023 02:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
2:00 189.9 F 4.2  
4:00 204.8 F 9.5  
6:00 221.6 F 7.1  
8:00 234.1 F 11.6  
10:00 236.0 F 10.0  
12:00 236.3 F 9.3  
14:00 251.8 F 16.9  
16:00 264.6 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.3 F 29.6  
2:00 271.9 F 29.7  
3:00 272.2 F 29.5  
4:00 271.6 F 29.5  
  
Min 271.4 F 29.1  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.2 F 0.0  
5:00 220.6 F 0.0  
10:00 234.8 F 0.0  
15:00 231.0 F 0.0  
20:00 227.0 F 0.0  
25:00 230.8 F 0.0  
30:00 227.4 F 0.0  
  
Drying Complete

FILLING: 2:25  
HEATING: 17:48  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:57:33  
CYCLE COMPLETE

05/16/2023 03:52 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_