
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005801
12:08 20/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:03
STERILIZING 3:58
136.4 C 222kPa 3:58
136.1 C 218kPa 4:28
136.4 C 224kPa 4:58
136.2 C 220kPa 5:28
136.2 C 222kPa 5:58
136.3 C 223kPa 6:28
136.4 C 223kPa 6:58
136.4 C 225kPa 7:28
136.4 C 225kPa 7:28
VENTING 7:28
AIR DRYING 8:21
CYCLE COMPLETE 12:05

Digital Signature #
CFB46C70CACBFF23
