
Midmark M11 - v1.0.0
Total Cycles: 3054

Name: RITTER 1
Sterilizer ID: V2178198
03/29/2023 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.0 F 0.0
2:00 122.6 F 0.0
4:00 147.4 F 1.7
6:00 174.7 F 5.2
8:00 198.4 F 10.0
10:00 217.1 F 5.5
12:00 225.0 F 7.9
14:00 235.0 F 12.5
16:00 235.2 F 9.3
18:00 238.5 F 9.7
20:00 246.5 F 13.6
22:00 258.6 F 20.2
24:00 269.7 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.6 F 29.2
2:00 271.8 F 28.4
3:00 272.3 F 29.2
4:00 272.4 F 29.0

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 197.3 F 0.0
10:00 212.0 F 0.0
15:00 212.1 F 0.0
20:00 215.1 F 0.0
25:00 217.7 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:02
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:03:09
CYCLE COMPLETE

03/29/2023 02:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____