
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001098
12:30 26/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:43
PRESSURIZING 4:43
STERILIZING 5:49
136.0 C 220kPa 5:49
136.4 C 225kPa 6:19
136.4 C 225kPa 6:49
136.4 C 224kPa 7:19
136.4 C 223kPa 7:49
136.0 C 222kPa 8:19
136.4 C 226kPa 8:49
136.4 C 224kPa 9:19
136.4 C 224kPa 9:19
VENTING 9:19
AIR DRYING 10:17
CYCLE COMPLETE 25:17

Digital Signature #
6EB7AB5B3794C093
