
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.1uS / 1.3ppm
CYCLE NUMBER 000021
16:39 13/01/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 4:05
STERILIZING 5:48
136.5 C 227kPa 5:48
136.5 C 228kPa 6:18
136.4 C 227kPa 6:48
136.4 C 227kPa 7:18
136.4 C 227kPa 7:48
136.5 C 226kPa 8:18
136.4 C 225kPa 8:48
136.4 C 227kPa 9:18
VENTING 9:18
AIR DRYING 10:34
CYCLE COMPLETE 40:34

Digital Signature #
1EF7D64DDF3390DC
