
Midmark M11 - v1.0.0
Total Cycles: 2953

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2023 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.1 F 0.0
2:00 191.7 F 3.7
4:00 197.6 F 8.0
6:00 219.4 F 13.5
8:00 229.3 F 8.7
10:00 239.1 F 10.8
12:00 241.1 F 12.0
14:00 240.6 F 10.7
16:00 254.3 F 17.7
18:00 266.4 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.2
2:00 271.5 F 28.4
3:00 272.0 F 29.1
4:00 272.2 F 29.0

Min 271.1 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 220.1 F 0.0
10:00 230.8 F 0.0
15:00 231.5 F 0.0
20:00 230.2 F 0.0
25:00 220.6 F 0.0
30:00 217.1 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

02/28/2023 11:56 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____